

## Scattering Meter Calibration Sheet

12/4/2013

Wavelength: 470

S/N BB2FLMBA-1133

Use the following equation to obtain "scaled" output values:

$$\beta(\theta_c) \text{ m}^{-1} \text{ sr}^{-1} = \text{Scale Factor} \times (\text{Output} - \text{Dark Counts})$$

• **Scale Factor for 470 nm** = 1.080E-05 (m<sup>-1</sup>sr<sup>-1</sup>)/counts

• **Output** = meter reading counts

• **Dark Counts** = 47 counts

Instrument Resolution = 1.2 counts 1.26E-05 (m<sup>-1</sup>sr<sup>-1</sup>)

### Definitions:

- **Scale Factor:** Calibration scale factor,  $\beta(\theta_c)$ /counts. Refer to User's Guide for derivation.
- **Output:** Measured signal output of the scattering meter.
- **Dark Counts:** Signal obtained by covering detector with black tape and submersing sensor in water.

Instrument Resolution: Standard deviation of 1 minute of collected data.